



**MAHARASHTRA NATURAL
GAS LIMITED**

**Calibration for RPD meter installed at MRS /
DRS at MNGL Pune.**

Date: 15.02.2018

Replies to Queries of the Bidders

SUB: Replies to queries of the bidders' for tender for calibration for RPD meter installed at MRS / DRS at MNGL Pune.

REF: Bid Document No. MNGL/C&P/2017-18/121 dated 01.02.2018

Dear Sir,

Following are the replies to queries of the bidders'. Kindly note the same.

Sr. No.	Bidder's Query / Clarification	MNGL's Reply
1		Additional Scope: Dismantling and re-installing of RPD meters will be in the scope of contractor. Any parts and fittings required will be provided by MNGL
2	Please provide us with the line size of the flowmeters to be calibrated.	Please refer the Annexure-I enclosed herewith.
3	Please confirm MNGL location ?. whether customer site or MNGL store. Also confirm the min quantity per LOT. Please confirm Meter transportation & Insurance from & to "MNGL to vendor store and back to MNGL store" Should be in MNGL scope.	There are different MNGL locations, each location details will be shared with you prior to the activity. Minimum Quantity per lot is 4 meters. Meter transportation / insurance will be in vendor scope..
4	Please confirm whether the laboratory should be NABL accredited or non NABL also accepted. Please consider Calibration / Testing in the scope as all utility accepting to both (technically both the laboratories are calculating meter error and pressure loss with compression of a reference standard meter. Utility IGL, ADANI, MGL, Gujrat Gas, SGL and other are widely accepting to both.	Laboratory should be NABL accredited. Both Calibration and Testing accepted but if RPD meter fails Testing then Calibration to be done with no extra cost.
5	Please note that RPD & Turbine meters must be transported and tested without OIL. However, OIL filling & top up shall be done at site after meter installation and same shall be in the scope of MNGL.	Accepted



**MAHARASHTRA NATURAL
GAS LIMITED**

**Calibration for RPD meter installed at MRS /
DRS at MNGL Pune.**

6	<p>There is no provision to issue the Test Certificate for 5years.</p> <p>However, to define the calibration frequency is purely authority of user.</p>	Accepted
7	<p>Minimum 10 working days' time require for Testing excluding transportation time.</p> <p>In case a planned periodic maintenance check or audit is going on for the laboratory the delivery lead time could be higher. However same shall be intimated to MNGL in well advance time. please confirm.</p>	Maximum 10 working days will be accepted apart from transportation time.
8	<p>Please confirm that the LD is not applicable for Service / Calibration</p>	Tender Conditions prevails
9	<p>Please extend the due date for at least 10 working days after receipt of query reply.</p>	Bid submission due date & time Extended upto 24.02.2018 @ 15.00 Hrs.
10	<p>As per scope of work, The Vendor shall transport the RPD meters from MNGL specified location to their calibration facility and back to MNGL location after calibration.</p> <p>We cannot arrange to collect & transport the meters from MNGL specified location to our calibration facility. Meters are required to be duly packed and dispatched at our facility by MNGL on freight paid basis. After calibration we shall dispatch back the meters to MNGL duly packed on freight to pay basis.-</p>	Tender Conditions Prevails
11	<p>Further, the NABL accreditation of laboratory must be valid throughout the period of contract. In case if NABL does not renew the accreditation given to laboratory after its present validity or terminates the accreditation in between the validity at any moment of time during the tenure of the contract, MNGL shall have all rights to terminate the contract</p>	Agreed

 MAHARASHTRA NATURAL GAS LIMITED	Calibration for RPD meter installed at MRS / DRS at MNGL Pune.
--	---

12	<p>Tender dose not state number of calibration points and flow 5 of these.</p> <p>Please confirm that the 5 point calibration to be done for flow ranges 10%, 25%, 50%, 75% & 100% of meter range</p>	Agreed
13	<p>Please confirm that bidder scope shall be only to calibrate the meters which are found mechanically & functionally OK and report actual errors. Any Repair / Servicing of the meters is not in scope of bidder</p>	Agreed
14	<p>Please confirm calibration contract period for successful bidder and also confirm minimum number of meters which would be sent for calibration during this period.</p>	<p>Period of the contract shall be till the completion of the activity.</p> <p>For Minimum quantity refer the Sr. No 03.</p>

Please submit the signed & stamped copy of Replies to Queries of the Bidders along with tender document with your techno-commercial offer as a token of acceptance.

All the others terms and conditions of the bid documents are unchanged.

With regards,

Shekhar D Kankrej
Dy. Manager (C&P)